New techniques

Disposable pudendal nerve stimulator: evaluation of the standard instrument and new device

J ROGERS, M M HENRY, AND J J MISIEWICZ

From the Department of Gastroenterology and Nutrition, Central Middlesex Hospital, London

SUMMARY A disposable version of the pudendal nerve stimulator using flexible printed circuit (FPC) technology has been developed and manufactured in our department. Evaluation of this instrument against the standard is reported.

In 1984 Kiff and Swash¹ were the first to describe pudendal nerve stimulation in investigation of the pathophysiology of idiopathic (neurogenic) faecal incontinence. Their technique was developed from that of electroejaculation described by Brindley,² for patients with impotence caused by paraplegia. The original device consisted of a rubber finger stall with two stimulating electrodes mounted on the tip and two recording electrodes mounted at the base. Though a fairly reliable instrument, it required disinfection in glutaraldehyde solution between patients and prolonged drying to prevent shorting of the electrodes.

A modified and more reliable version of this reusable type of pudendal nerve stimulator was developed in our department and is used by a number of anorectal physiology units in this country and overseas. The main drawback with this device is that it needs thorough cleaning between subjects, to prevent cross infection. Using the techniques used in the manufacture of flexible printed circuits we have developed a disposable pudendal nerve stimulator of the design shown in Figure 1. The electrodes are mounted in the same relative positions as in the reusable model and are of the same dimensions. The flexible printed circuits terminals from the new stimulator are connected to the EMG machine by a reusable lead by means of a quick release clip, which can be engaged and disengaged easily by a single

Fig. 1 New disposable (top) and standard reusable: pudendal nerve stimulator.

hand operation. The connections are automatically engaged when the clip is closed. The stimulator has a self adhesive backing and will securely fix to most types of latex gloves, conforming easily to the contours of the index finger. The author has found that the best method is to fix the stimulator to a standard disposable glove worn over a latex examination glove for ease of operation, disposal and comfort for the patient. The stimulator is inserted and used in the same way as the reusable type.^{1,3,4}

Methods

ASSESSMENT

The new stimulator has been evaluated during

Address for correspondence: Mr J Rogers, Dept of Gastroenterology and Nutrition, Central Middlesex Hospital, Acton Lane, London NW10 7NS.

Received for publication 25 February 1988.

Table 1 *Reusable v disposable stimulator*

	<i>Reusable</i>	<i>Disposable</i>
Mean PNTML ms (SD)	1.98 (0.25)	1.96 (0.25)
Mean difference (SD)	0.02 (0.1)	
95% CI bias	from -0.05 to 0.08	
95% CI upper limit*	from 0.18 to 0.25	
95% CI lower limit*	from -0.21 to -0.15	

*Limits of agreement set at 2SD. PNTML=pudal nerve terminal motor latencies.

routine anorectal physiology studies in 12 subjects. Six had paired measurements of their right and left pudendal nerve terminal motor latencies using both instruments in a randomised order. In the other six the measurements were repeated for both stimulators to assess the repeatability of each method. In all studies the stimulus strength and duration were constant. A supramaximal 50 v stimulus of 0.1 ms duration at a rate of 1 pps was used to stimulate the pudendal nerve at the level of the ischial spines. Serial recordings made on a Medelec MS6 EMG machine. Twelve paired results were obtained for each part of the study. The statistical method for assessing agreement between the two methods and repeatability for each method was taken from Bland and Altman.⁵

Results

AGREEMENT BETWEEN THE TWO METHODS
(Table 1)

Figure 2 shows the results of the difference in pudendal nerve terminal motor latencies between the old and the new stimulator plotted against the average pudendal nerve terminal motor latencies by both techniques. The results show the mean differ-

Fig. 2 Results of test of agreement: reusable v disposable stimulator.

ence (bias), the upper and lower limits of agreement ((2SD) from the mean difference) and their corresponding 95% confidence intervals. The results show a small bias between the two techniques but the 95% confidence intervals encompass zero and there is no statistical difference between the results obtained by the two techniques. The limits of agreement are estimated at (2SD) as this is where we would expect 95% of the differences to lie assuming a normal gaussian distribution of the data. The results are interesting in that the range of these limits is in the order of 0.4 ms which is large considering the range of latencies quoted in the literature for different patient groups.¹⁴⁶ We must be aware, however, that this range of limits is for agreement between the two methods under comparison. The accuracy of the techniques themselves are assessed by their repeatability.

REPEATABILITY OF EACH TECHNIQUE (Table 2)
Figure 3a which shows the results of repeatability for the reusable stimulator and Figure 3b which shows

Fig. 3 Results of repeatability of stimulators. a Reusable stimulator, b Disposable stimulator.

Fig. 4 Two tracings of the compound motor action potential of the external anal sphincter taken from the same subject after stimulation of the left pudendal nerve by reusable (top) and disposable stimulator.

the results of repeatability for the disposable stimulator. The limits of agreement are similar for the two techniques and the range is at the most 0.34 ms but while there is no bias in the mean difference by repeated measurement with the new stimulator there is a minute negative bias in the mean difference recorded by the standard stimulator. The repeatability coefficient⁵ for the new stimulator is 0.11 ms and for the standard is 0.12 ms.

Discussion

A disposable pudendal nerve stimulator has the advantages of ease of operation, smaller instrument diameter, and no risk of transmission of infection between subjects. The drawback is that the flat electrode surfaces are more difficult to apply and 'skating' of the stimulus artefact across the surface of the device sometimes occurs producing an easily recognisable false potential. This is remedied by removing the stimulator and wiping any excess lubricating gel or mucus from the recording surfaces.

Table 2 Repeatability of reusable and disposable stimulator

Reusable:	First sample	Second sample
Mean PNTML ms (SD)	1.95 (0.17)	1.95 (0.14)
Mean difference (SD)	0.008 (0.06)	
95% CI bias	from -0.05 to 0.03	
95% CI upper limit*	from 0.08 to 0.16	
95% CI lower limit*	from -0.17 to -0.09	
Disposable:	First sample	Second sample
Mean PNTML ms (SD)	1.93 (0.14)	1.93 (0.14)
Mean difference	$1.8e^{-17}$	
95% CI bias	-0.3 to 0.03	
95% CI upper limit*	from 0.07 to 0.15	
95% CI lower limit*	from -0.15 to -0.07	

*Limits of agreement set at 2SD.

Stimulus artifact occurs less often using the reusable stimulator as the electrode surfaces are more easily applied and are better isolated. The bias between the two techniques is small and the limits of agreement are probably wide enough to allow the control data recorded by one instrument to be applied to the other. The only reservation to this statement is that this original assessment has been made on a relatively small subject group, but the precision of the estimate as shown by the 95% CI is good.

AVAILABILITY

Both types of stimulator are available and may be obtained from The Department of Gastroenterology and Nutrition, Central Middlesex Hospital, Acton Lane, London NW10 7NS.

The authors wish to thank Mr Alan Pitts for his technical assistance.

References

- 1 Kiff ES, Swash M. Slowed conduction in the pudendal nerves in idiopathic (neurogenic) faecal incontinence. *Br J Surg* 1987; **71**: 614-6.
- 2 Brindley GS. Electroejaculation: its technique, neurological implications and uses. *J Neurol Neurosurg Psychiatry* 1981; **44**: 9-18.
- 3 Henry MM, Snooks SJ, Barnes PRH, Swash M. Investigation of disorders of the anorectum and colon. *Ann R Coll Surg Engl* 1985; **67**: 354-60.
- 4 Rogers J, Henry MM, Misiewicz JJ. Combined sensory and motor deficit in primary neuropathic faecal incontinence. *Gut* 1988; **29**: 5-9.
- 5 Bland JM, Altman DG. Statistical methods for assessing agreement between two methods of clinical measurement. *Lancet* 1986; **i**: 307-10.
- 6 Kiff ES, Swash M. Normal proximal and delayed distal conduction in the pudendal nerves of patients with idiopathic (neurogenic) faecal incontinence. *J Neurol Neurosurg Psychiatry* 1984; **47**: 820-3.