A new variable stiffness colonoscope makes colonoscopy easier: a randomised controlled trial

J C Brooker, B P Saunders, S G Shah, C B Williams

Abstract

Background—Colonoscopy remains technically difficult in 10–20% of procedures due to variable colonic anatomy and fixation. The ability to vary endoscope shaft flexibility may help insertion to the caecum.

Methods—Consecutive patients attending for day case colonoscopy were randomised to examination with either the conventional Olympus CF200HL (200HL) or a new variable stiffness (VS) colonoscope. Intubation time, use of stiffening function, and patient pain scores were compared.

Results—Of 100 cases, 43 were performed with the 200HL and 57 with the VS. Four incomplete examinations occurred with the 200HL (two sigmoid fixations, two benign strictures) and two with the VS (one obstructing cancer, one fixed sigmoid). Changing to the paediatric scope was successful in all but one patient from each group (obstructive lesions). Stiff mode was applied 23 times in 18 patients and was effective in 15 of these. Intubation time was quicker with the VS (median 6 minutes 32 seconds) than with the 200HL (median 10 minutes 35 seconds) ($p=0.0005$). Pain scores were less with the VS (median 7) than with the 200HL (median 24) ($p=0.0081$).

Conclusions—The variable stiffness colonoscope combines paediatric shaft characteristics with the ability to stiffen when needed. This instrument significantly reduces intubation time and patient discomfort. Further comparisons should be made with the newest colonoscopes which are less stiff.

(Gut 2000;46:801–805)

Keywords: colonoscopy; colonoscopes; technology; pain

The Wolfson Unit for Endoscopy, St Mark's Hospital, Northwick Park, London HA1 3UJ, UK

J C Brooker
B P Saunders
S G Shah
C B Williams

Correspondence to:
Dr B Saunders. Email:
wolfsonendoscopy@btinternet.com

Accepted for publication
11 January 2000

ably by reducing the amount of stretch caused by looping.⁴

Several methods have been proposed for varying colonoscope shaft stiffness. An overtube⁵ is often successful in controlling sigmoid colon looping but cannot prevent proximal loops, is cumbersome to use, and may cause complications.⁶ Success has been reported with stiffening wires inserted through the biopsy channel^{7–9} but their use prolongs the examination, disables the suction/instrumentation channel, and can also damage the colonoscope.¹⁰ A method is also described for exchanging between paediatric and conventional instruments over a guidewire, after

Figure 1 The control head of the variable stiffness colonoscope (Olympus XCF-SH230L). The dial for controlling the stiffness is arrowed. Specifications: distal end outer diameter 13.6 mm, flexible tube outer diameter 12.9 mm, instrument channel 3.2 mm, and working length 168 cm.

Figure 2 The stiffness of three colonoscope shafts: the conventional colonoscope (200HL), the variable stiffness colonoscope in paediatric mode (VS paed) and at maximum stiffness (VS stiff), and the paediatric colonoscope. Stiffness was measured at 10 cm intervals along the shaft from the 20 cm mark using a tension gauge to depress each point by 5 cm when suspended across a 20 cm gap. Each value is the mean of three readings.

Abbreviations used in this paper: VS colonoscope, variable stiffness colonoscope; MEI, magnetic endoscope imaging.

Table 1 Patient characteristics in the groups randomised to be examined with the variable stiffness (VS) and conventional (200HL) colonoscopes

	VS (n=57)	200HL (n=43)
Age (median (range))‡	58 (26–82)	55 (24–86)
Female*	33	29
Male*	24	14
Anxiety (median (range))†	31 (0–95)	29 (0–100)
Hysterectomy††	1	3
Diverticular disease**	8	11

‡p=0.65 (Mann-Whitney); *p=0.4 (Fisher's exact); †p=0.95 (Mann-Whitney); ††p=0.3116 (Fisher's exact); **p=0.1985 (Fisher's exact).

Table 2 Indications for colonoscopy in the groups examined with the variable stiffness (VS) and conventional (200HL) colonoscopes

	VS (n=57)	200HL (n=43)	p (Fisher's exact test)
IBD	22	10	0.1311
Bowel symptoms	20	20	0.3041
Cancer/polyp screening	15	13	0.8223

IBD, assessment of extent of inflammatory bowel disease, extensive ulcerative colitis surveillance, or differential diagnosis. Bowel symptoms, change in bowel habit, rectal bleeding, abdominal pain, anaemia, or abnormal barium. Cancer/polyp screening, family history of colon cancer or past history of polyps or cancer.

passage through the sigmoid,³ but this is time consuming and requires fluoroscopy.

Our hypothesis was that a new colonoscope with in-built variable shaft stiffness should improve intubation time and reduce patient discomfort during insertion (by virtue of increased flexibility in the paediatric setting), without adversely affecting completion rates (because of the ability to avoid reformation of loops after straightening by increasing the shaft stiffness).

Materials and methods

VARIABLE STIFFNESS (VS) COLONOSCOPE

The prototype colonoscope (Olympus XCF-SH230L) has a graduated dial situated just below the control head of the colonoscope (fig 1) which can be turned manually to vary shaft stiffness from minimum to maximum. In all other respects it is a normal video colonoscope (shaft diameter 12.9 mm, working length 168 cm, instrument channel 3.2 mm). The mechanism of adjustable shaft stiffness is similar to that used in commercially available "through the channel" stiffening wires. These have a central cable with a surrounding metal helical coil. Tension applied to the cable compresses and stiffens the helix and colonoscope, whereas loosening has the opposite effect resulting in paediatric characteristics. The stiffening mechanism terminates 30 cm from the instrument

tip so that the distal portion remains relatively floppy. For the purposes of this study minimum stiffness was used as the default setting. When greater shaft stiffness was required the maximum setting was used.

CONTROL COLONOSCOPE

The control instrument used in this study was the standard Olympus CF200HL (200HL): shaft diameter 13.3 mm, instrument channel 3.2 mm, working length 165 cm. This colonoscope was selected because it was the most widely used in UK endoscopy departments in early 1998 (personal communications, Olympus Keymed, UK) when the study began.

The Olympus PCF 230 paediatric colonoscope (shaft diameter 11.3 mm, instrument channel 3.2 mm, working length 130 cm) was used to attempt completion in cases that had failed due to fixation or stricture.

The shaft stiffness of the variable stiffness, 200HL, and PCF colonoscopes was measured for comparison. A tension gauge was used at 10 cm intervals to determine the force required (mean of three readings) to depress the shaft by 5 cm when suspended between two points 20 cm apart (fig 2).

PATIENTS

Consecutive patients attending for routine day case colonoscopy were randomised (random number generator) to be examined with either the 200HL or VS colonoscope. Patients and admission/recovery nurses were blinded to the type of colonoscope used. Patients also scheduled for gastroscopy or those with previous colonic resection were excluded. All patients were given routine bowel preparation: 24 hour dietary restriction, magnesium citrate (29.5 g sachets×2) and senna granules (13 g weight). Pre-procedure anxiety was assessed with a 100 mm visual analogue scale (0=not worried, 100=extremely anxious) completed by the patient after explanation by the admitting nurse. Patients who requested sedation were given standard initial intravenous medication with midazolam 1.25 mg, pethidine 25 mg, and hyoscine 20 mg. Additional doses were given as required. Each procedure was performed by one of two experienced colonoscopists (CBW and BPS). All examinations were commenced in the left lateral position. Caecal intubation time was recorded as the interval between initial insertion through the anal canal and intubation of the caecal pole, assessed by recognition of the appendix orifice and ileo-caecal valve. Cases in which caecal intubation failed with the allocated instrument were counted as failed examinations. Data from these cases were excluded from the analysis of intubation time and pain. The paediatric colonoscope was used to attempt completion in those cases that failed with either of the randomised instruments.

The maximum stiffness setting of the VS scope was used only when shaft looping occurred at the default (minimum stiffness) setting and did not respond to abdominal hand pressure or change in position. Maximum stiffness was only applied after fully straightening

Table 3 Incomplete examinations. Cases in which caecal intubation with the designated colonoscope was unsuccessful, listed in order of occurrence. The colonoscope used, reason for failure, and result of converting to the paediatric colonoscope are also given

Case	Colonoscope*	Reason for failure	Completed with paediatric colonoscope
31	VS	Obstructing sigmoid cancer	No
57	200HL	Fixed sigmoid angulation	Yes
70	VS	Fixed sigmoid (diverticular disease)	Yes
81	200HL	NSAID induced diaphragm (stricture) transverse colon	No
89	200HL	Fixed sigmoid (diverticular disease)	Yes
95	200HL	Rectosigmoid stricture (Crohn's)	Yes

*p=0.3979 (Fisher's exact).

Figure 3 Caecal intubation times and median values for examinations completed with the variable stiffness (VS) and conventional (200HL) colonoscopes. *Mann Whitney U test.

Figure 4 Pain scores assessed using a 100 mm visual analogue scale for examinations completed with the variable stiffness (VS) and conventional (200HL) colonoscopes. *Mann Whitney U test.

Figure 5 Efficacy of stiffening the colonoscope shaft in overcoming recurrent looping according to the estimated location of the tip.

Table 4 Male and female pain scores. Patient pain scores (in examinations that were completed with the assigned instrument) for males and females for the variable stiffness (VS) and control (200HL) colonoscopes

	Male (range) [n]	Female (range) [n]	p Value*
200HL	8 (0–70) [13]	39 (1–85) [27]	0.0122
VS	5.5 (0–48) [23]	9 (2–82) [31]	0.3832
p Value*	0.9119	0.0022	

*Mann-Whitney U test.

Table 5 Male and female caecal intubation times. Caecal intubation times (in examinations that were completed with the assigned instrument) for males with females for the variable stiffness (VS) and control (200HL) colonoscopes

	Male (range) [n]	Female (range) [n]	p Value*
200HL	07:40 (3:50–13:40) [13]	11:45 (4:33–22:33) [27]	0.0136
VS	4:48 (1:55–19:40) [23]	7:26 (3:07–18:14) [31]	0.0038
p Value*	0.0395	0.0038	

Times are in minutes:seconds.

*Mann-Whitney U test.

the instrument shaft. Whenever shaft stiffening was used, colonoscope tip location was estimated as well as the efficacy of the manoeuvre. A successful result was considered to be immediate advancement of the colonoscope tip. After the manoeuvre the VS colonoscope was returned to minimum stiffness. A 100 mm visual analogue scale for pain (0=painless, 100=extremely painful) during colonoscopy was completed by the patients immediately prior to discharge, usually within 45 minutes of completing the examination.

STATISTICAL ANALYSIS

Differences in anxiety score, pain score, and intubation time were compared using the Mann-Whitney U test. The characteristics of the two groups of patients were compared using Fisher's exact test. All p values were two-tailed and $p < 0.05$ was considered significant.

Results

One hundred consecutive eligible patients were randomised. There was no significant difference between the two patient groups in age, sex, anxiety, previous diverticular disease, or hysterectomy (table 1). There was also no difference in indications for colonoscopy (table 2) or amount of sedation used (VS group: median pethidine dose 25 mg (range 0–75), median midazolam dose 1.5 mg (0–3); 200HL group: median pethidine dose 37.5 mg (0–100), median midazolam dose 1.5 mg (0–2.5)). Fifty five cases were completed with the VS and 39/43 with the 200HL colonoscopes. There were no complications in either group.

The six incomplete examinations are detailed in table 3. Four were the result of benign fixation of the sigmoid (two diverticular disease, one Crohn's disease, one no cause stated). All were completed after changing to the paediatric colonoscope. The other two failures, an obstructing sigmoid cancer and a tight NSAID induced transverse colon stricture, could not be passed even with the paediatric colonoscope.

Of the completed examinations, intubation time was significantly quicker ($p = 0.0005$) using the VS colonoscope (fig 3) (VS: median time 6 minutes 32 seconds (range 1:50–19:35); 200HL: median time 10 minutes 35 seconds (3:45–22:35)). Patient pain was significantly less ($p = 0.0081$) with the VS colonoscope (fig 4) (VS: median pain score 7 (range 0–82); 200HL: median pain score 24 (0–85)). During the examinations performed with the VS scope, shaft stiffness was increased because of uncontrollable looping 23 times in 18 patients and was considered to be effective on 15 of 23 occasions (fig 5). Analysis by patient sex is detailed in tables 4 and 5. Women recorded significantly more pain than men with the 200HL but there was no difference between the sexes when the VS colonoscope was used. Intubation times were quicker in men than women for both instruments, and quicker in women with the VS colonoscope than with the 200HL.

Discussion

A fixed sigmoid colon or recurrent looping are the commonest difficulties that arise during colonoscopy insertion and may coexist in the same patient.¹ The flexibility and narrow calibre of a paediatric colonoscope enables an easier, less traumatic passage through a narrowed or angulated colon.²⁻⁴ However, the tendency of the paediatric instrument to form loops and its limited length (130 cm) compared with a standard colonoscope (160 cm) reduces the frequency of caecal intubation.³ In most cases the anatomy and characteristics of a patient's colon are not known before starting the procedure. It is only when difficulty occurs or the patient experiences excessive pain with a standard colonoscope that the need to consider changing to the paediatric instrument becomes apparent.

In the patient with a redundant or mobile colon, recurrent looping commonly occurs even with the stiffer conventional adult colonoscope. Devices for increasing shaft stiffness without the need for a change of instrument have included a wire inserted through the operating channel⁷⁻⁹ and the "splinting" overtube¹¹ or its split version.⁵ Although effective, both of these approaches have drawbacks. Stiffening wires obstruct the suction/biopsy channel thus impairing performance. They may also damage the colonoscope.¹⁰ An overtube is cumbersome, often requires fluoroscopy for effective use, may cause complications such as perforation,¹¹ and even if successful in the sigmoid colon it does not prevent proximal looping.

In this study we have evaluated a new colonoscope which combines the shaft flexibility of a paediatric colonoscope with the working length of an adult colonoscope and a built in mechanism for stiffening the shaft when looping occurs. Thus the shaft characteristics of the VS colonoscope can be adapted to suit the fixed or looping colon without wasting time or compromising the procedure and with no risk of damaging the patient or instrument.

We have shown that the new VS colonoscope significantly reduces intubation time and patient discomfort compared with a standard adult instrument. The reduction in intubation time was probably the result of two main factors. Firstly, being more flexible, the prototype instrument traversed the sigmoid colon more easily. Secondly, when looping did not respond to hand pressure or position change, straightening the shaft and increasing the stiffness resulted in better progress on 15 of 23 occasions. Because the endoscopists could not be blinded, it is possible that they unintentionally tried harder with the new instrument. However, it is unlikely that this bias alone could have accounted for the magnitude of difference that we observed between the two scopes.

The reduction in pain score recorded by the patient was probably the result of less stretching of sigmoid colon loops by the floppy shaft of the VS colonoscope when in paediatric mode. The shorter procedure time may also have contributed. A much larger study is needed to determine whether reducing the

stretching forces exerted on the sigmoid colon and attachments also lessen the risk of perforation during diagnostic colonoscopy with this instrument. Although we showed no difference in sedation requirements between the two groups, lessening the trauma of insertion through the sigmoid may reduce, and in some cases eliminate, the need for medication. This should encourage the increasingly popular practice of non-sedated colonoscopy¹² with a consequent reduction of cardiorespiratory complications.

It is well documented¹¹ that colonoscopy is more difficult, takes longer, and is less well tolerated in women than men. This is probably because of the longer female colon,¹¹⁻¹³ increased incidence of pelvic adhesions, and perhaps greater sensitivity to pain in the sigmoid colon. The results of analysis by patient sex are therefore of interest. We have shown that, with the new instrument, intubation times are reduced by an average of 3-4 minutes in both men and women, and that pain scores in females are reduced to levels experienced by men.

Caecal intubation was not achieved with the designated colonoscope in six cases, two with the VS and four with the 200HL colonoscope. Only one case in each group, caused by obstructing lesions, could not be completed after changing to a paediatric colonoscope. The numbers in this study are insufficient to show a significant difference in completion between the two instruments. However, it is logical that the floppier VS colonoscope should be able to negotiate tight angulations or fixed loops more easily. On one occasion in this study, a fixed sigmoid could only be successfully intubated after exchanging the VS for a narrower, floppier paediatric colonoscope. It might therefore be argued that a modified VS colonoscope, with a similar calibre to that of the paediatric instrument, if used routinely would obviate the need to change instruments for a very narrowed or fixed sigmoid colon. However, in pilot studies we have found that difficulty gripping the narrow shaft of such an instrument makes routine use less practical. Thus the more flexible, narrow calibred, paediatric colonoscope with its unique bending section characteristics is likely to remain an important back up for the occasional very narrowed or fixed sigmoid colon.

We have shown that the stiffening function of the VS scope was effective 15/23 times overall, with varying degrees of success at different locations. It was used most often at locations where looping most commonly occurs—that is, the sigmoid colon, splenic flexure, and transverse colon—where it was effective on 2/4, 4/4, and 4/9 occasions, respectively. The effectiveness of increasing shaft stiffness will probably improve with practice as endoscopists become adept at identifying the circumstances in which it will help insertion. We have previously shown¹⁴ that even experienced endoscopists are frequently inaccurate when estimating colonoscopy location and loop configuration. This might confound the utility of the variable stiffness function in some cases. However, new

developments in imaging technology, such as magnetic endoscope imaging (MEI)¹⁵ which shows the colonoscope configuration in three dimensions, will enable the user to determine the precise nature and location of a loop. The addition of MEI may therefore further enhance the effect of adjusting shaft stiffness, allowing optimal straightening of the shaft before increased stiffness is applied. The availability, in the near future, of colonoscopes that integrate "variable stiffness" and built in MEI sensor coils, will allow these two features to be evaluated simultaneously.

For safety reasons the developers of the VS scope have limited the maximum stiffness available. Our own measurements have shown that even at the highest setting the stiffness is less than that of the CF200HL. In our experience even the relatively stiff conventional adult colonoscope is frequently not stiff enough to prevent looping. The design of the VS colonoscope will therefore probably need to be modified to increase the range of stiffness available.

As colonoscopic surveillance and population screening for colorectal cancer gain wider acceptance¹⁶ the number of colonoscopies performed on asymptomatic, healthy individuals will rise. This will create even more demand for quick, accurate, comfortable, and safe colonoscopy. We believe that this innovative instrument will be a valuable tool in achieving these aims and will probably become the colonoscope of first choice. Further studies will determine if the VS colonoscope confers the same or even more pronounced advantages when used by non-experts.

In conclusion, the new variable stiffness colonoscope improves the performance of colonoscopy in expert hands by reducing intubation time and patient discomfort. Modifications that may enhance the efficacy further

include more floppiness in the paediatric setting and greater stiffness at the maximum setting. Further comparison with the newer designs of conventional colonoscope, which have more flexible shafts, will be of interest. Future studies should examine the usefulness of the VS colonoscope in training, in non-sedated colonoscopy, and in combination with new imaging techniques such as magnetic endoscope imaging.

- 1 Saunders BP, Macrae F, Williams CB. What makes colonoscopy difficult? *Gut* 1993;34:A181.
- 2 Bat L, Williams CB. Usefulness of paediatric colonoscopes in adult colonoscopy. *Gastrointest Endosc* 1989;35:329-32.
- 3 Ness RM, Gottlieb K, Rex DK, Lehman GA. Difficult sigmoid colon intubation: guidewire exchange technique. *Gastrointest Endosc* 1996;44:99-101.
- 4 Rogers BHG. The use of small calibre endoscopes in selected cases increases the success rate of colonoscopy. *Gastrointest Endosc* 1989;35:352.
- 5 Williams CB. Split overtube for easier colonoscopy. *Gastrointest Endosc* 1983;24:A492.
- 6 Shinya H, Wolff WL. Colonoscopy. *Surg Annu* 1976;86:257-95.
- 7 Kitano A, Okawa K, Obata A, et al. The leading cord method of colonofibrescopy. *Dis Colon Rectum* 1986;29:882-4.
- 8 Deyhle P, Demling L. Colonoscopy—technique, results, indications. *Endoscopy* 1971;3:143-51.
- 9 Sullivan MJ. Variable stiffening device for colonoscopy. *Gastrointest Endosc* 1990;36:642-3.
- 10 Ruffolo TA, Lehman GA, Rex D. Colonoscope damage from internal straightener use. *Gastrointest Endosc* 1991;37:107-8.
- 11 Saunders BP, Fukumoto M, Halligan S, et al. Why is colonoscopy more difficult in women? *Gastrointest Endosc* 1996;43:124-6.
- 12 Rex D, Imperiale TF, Portish V. Patients willing to try colonoscopy without sedation: associated clinical factors and results of a randomised controlled trial. *Gastrointest Endosc* 1999;49:554-9.
- 13 Hull T, Church JM. Colonoscopy—how difficult, how painful? *Surg Endosc* 1994;8:784-7.
- 14 Shah SG, Saunders BP, Brooker JC, Williams CB. What happens during routine colonoscopy? An audit using magnetic positional imaging (MPI). *Gastrointest Endosc* 1999;49:AB175.
- 15 Saunders BP, Bladen J, Williams CB, Bell GD. First clinical results with a real-time electronic imager as an aid to colonoscopy. *Gut* 1995;36:913-17.
- 16 Nelson DB, McQuaid KR, Bond JH, Macrae F. Population-based colonoscopy screening for colorectal cancer is feasible and safe: preliminary results from the VA colonoscopy screening trial. *Gastrointest Endosc* 1999;49:AB65.